



Product Service

Attestation of Conformity

No. N8A 111395 0005 Rev. 00

Holder of Certificate: **Baoding Longer Precision Pump Co., Ltd**
4th Floor, Area B, Building 6
National University Science Park
071051 Baoding, Hebei
PEOPLE'S REPUBLIC OF CHINA

Product: **Control Equipment
(Peristaltic Pump)**

Model(s): **BT100-3J**

Parameters: Rated Input Voltage: 100-240Va.c.
Rated frequency: 50/60Hz
Rated Current / Power: 0.56A
IP degree: IP 20

Tested according to: EN 61010-1:2010/A1:2019
EN IEC 61010-2-201:2018

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: www.tuvsud.com/ps-cert

Test report no.: 64105213017701

Date, 2021-07-26

(Frank Zhu)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.